

**Met One Instruments, Inc.**

3206 Main Street, Suite 106, Rowlett, TX 75088

Phone: (972) 412-4715; Fax: (972) 412-4716

**Test Data ADA Wind Direction Alignment Fixture**Model No: 3269Serial Number: R12024Customer: USEPASales Order: 128687Customer P.O.: VISA

Standard used:

DMM

Fluke 287

Serial No.: 40930020 Calibration Due: 1/23/2019

Alignment (Degrees)	ADA Reading (Degrees)	Error (Degrees)	Spec (Degrees)
0/360	1.1	1.1	+/- 3
10	10.2	0.2	+/- 3
90	89.9	-0.1	+/- 3
180	179.8	-0.2	+/- 3
270	269.6	-0.4	+/- 3
350	349.2	-0.8	+/- 3

Technician: Dan FitchDate: 4/5/2018Inspected by: Byron DanksDate: 5/3/18

The standards used for calibration have accuracies equal to or greater than the instruments tested. These standards are on record and are traceable to NIST to the extent allowed by the institute's calibration facility. Unless otherwise stated hereon, all instruments are calibrated to meet the manufacturer's published specifications. The calibration system complies with MIL-STD-45662A